


## HANDLING INSTRUCTIONS

### EE894 - Digital Measurement Module for CO<sub>2</sub>, Relative Humidity, Temperature and Barometric Pressure





#### CAUTION

- EE894 is an ESD sensitive devices and shall be handled with corresponding precautions.
- EE894 and mainly the CO<sub>2</sub> sensing cell shall not be exposed to any mechanical stress during installation or operation. Mechanical stress on the sensing cell can lead to relevant measurement errors.

CO<sub>2</sub> Sensing Cell



#### HANDLING AND MOUNTING



#### TECHNICAL DATA AND SUPPORT LITERATURE

Please see data sheet, specification, application notes and tutorials at:

- [www.epluse.com/ee894](http://www.epluse.com/ee894)

## INFORMATION

+43 7235 605 0 / [info@epluse.com](mailto:info@epluse.com)

Langwiesen 7 • A-4209 Engerwitzdorf  
Tel: +43 7235 605-0 • Fax: +43 7235 605-8  
[info@epluse.com](mailto:info@epluse.com) • [www.epluse.com](http://www.epluse.com)

LG Linz Fn 165761 t • UID-Nr. ATU44043101  
Place of Jurisdiction: A-4020 Linz • DVR0962759

